



## **ISO/IEC 17025**

Pyranometer & Pyrhelimeter Calibration Procedures

Rev 2: May 2021



**EKO Instruments Co. Ltd. Calibration Laboratory** is accredited to perform calibration of pyranometers and pyrhemometers in accordance with **ISO/IEC 17025:2017**.



**EKO INSTRUMENTS CO.,LTD.**  
1-21-8 Hatagaya, Shibuya-ku,  
Tokyo 151-0072 Japan  
P. +81.3.3469.6713  
F. +81.3.3469.6719  
www.eko.co.jp

## Solar Sensor Calibration Procedures

At EKO Instruments Co. Ltd.

### ISO/IEC 17025:2017



EKO Instruments Co., Ltd. (EKO) calibration laboratory is accredited and certified by PJLA (Ref: **#74158**) to perform pyranometer and pyrhemometer calibrations per the requirements of **ISO/IEC 17025:2017**, which are relevant to calibration and testing.

EKO offers a manufacturer calibration service for pyranometers and pyrhemometers in-house. Based on the applied calibration methods, EKO exhibits the highest level of technical and managerial expertise for solar sensor calibrations compliant to the international standards defined by **ISO/IEC 17025:2017** for ISO9847 (Indoor method) and ISO9059 (Outdoor method), as defined by the scope of calibrations Appendix 1.

**ISO/IEC 17025:2017** provides a globally accepted basis for laboratory accreditation that specifies the most stringent management and technical requirements. With calibrations performed at the EKO Instruments laboratory, we enable our customers to:

- Clearly identify the applied calibration methods and precision.
- Be traceable to the World Radiation Reference (WRR) through defined industrial standards:
  - ISO9846 Calibration of a pyranometer using a pyrhemometer
  - ISO9847 Calibration of field pyranometer by comparison to a reference pyranometer
  - ISO9059 Calibration of field pyrhemometers by comparison to a reference pyrhemometer
- Obtain repeatable and reliable calibration test results through consistent operations yearly examined independently to assure strict compliance to the requirements.

Customers will obtain the highest level of confidence when purchasing an **ISO/IEC 17025:2017** calibrated sensor. EKO's accredited laboratory is regularly re-examined to ensure that they maintain their standards of technical expertise.



## CONTENTS

|  |    |
|--|----|
| <b>CALIBRATION</b> .....                         | 4  |
| Calibration Traceability.....                    | 4  |
| Reference Sensor Units.....                      | 5  |
| Calibration Procedures.....                      | 6  |
| Calibrating a Field Pyrheliometer.....           | 8  |
| Calibrating a Field Pyranometer.....             | 11 |
| <b>APPENDIX</b> .....                            | 13 |
| 1. Scope of ISO/IEC 17025:2017 Calibrations..... | 13 |
| 2. Solar Irradiance Definitions.....             | 14 |
| 3. Measurement Devices.....                      | 15 |
| 4. Criteria and Environmental Conditions.....    | 16 |
| 5. Calibration Certificates.....                 | 17 |
| 6. Measurement Uncertainty.....                  | 18 |
| 7. Sensor Characteristics.....                   | 19 |

## CALIBRATION


Calibration is the process of comparing the output signal of a sensor with the standard value in a defined calibration environment. Calibration is necessary to establish the correct value, and proper accuracy, of a sensor.

For pyranometers or pyrhemimeters, we aim to determine the correlation between the detector response (e.g.,  $\mu\text{V}$ ) and irradiance ( $\text{W}/\text{m}^2$ ), known as sensitivity ( $\mu\text{V}/(\text{W}/\text{m}^2)$ ).

The following section provides a summary of the calibration methods and definitions of terms used.

### Calibration Traceability

According to the traceability flow chart below, EKO pyranometers and pyrhemimeters are calibrated and traceable to the 'World Radiometric Reference' (WRR).





## Reference Sensor Units

Reference sensor units used for calibration:

- [Absolute Cavity Radiometer PMO6](#): This unit belongs to EKO and is traceable to the WRR maintained at the World Radiation Centre (WRC) in Davos, Switzerland. The PMO6 is categorised under ISO 9060:2018 as a Class AA pyr heliometer.
- Every 5 years, the PMO6 is calibrated against the WRR during the International Pyr heliometer Comparison (IPC) to maintain its reference status. The WRR is based on a cluster of 6 models of Absolute Cavity Radiometers, referred to as the World Standard Group (WSG).
- [EKO's MS-57 reference Pyr heliometers](#) are calibrated directly against the EKO PMO6, according to the standard method defined by 'ISO 9059:1990, Solar energy -- Calibration of field pyr heliometers by comparison to a reference pyr heliometer'. These reference pyr heliometers are calibrated once per year.
- [Pyr heliometer products](#) are calibrated according to the standard method defined by 'ISO 9059:1990 Solar energy -- Calibration of field pyr heliometers by comparison to a reference pyr heliometer'. This involves comparing the sensor against a reference Pyr heliometer, using measurements taken outdoors simultaneously.
- [Reference Pyranometer](#) units are established for each product model and calibrated according to the standard method defined by 'ISO 9846:1993 Solar energy -- Calibration of a pyranometer using a pyr heliometer'. These sensors are compared against the PMO6, using measurements taken outdoors simultaneously. Reference pyranometers are calibrated once per year.
- [Pyranometer products](#) are calibrated according to the standard method defined by 'ISO 9847:1992 Solar energy -- Calibration of field pyranometers by comparison to a reference pyranometer', with measurements taken indoors.

## Calibration Procedures

### Reference Pyrheliometer Unit

An EKO MS-57 Pyrheliometer is used as a reference for outdoor calibrations of pyrheliometers.

This unit is itself calibrated through outdoor measurements and compared against the Absolute Cavity Radiometer PMO6; per the calibration method defined by 'ISO 9059:1990 Solar energy -- Calibration of field pyrheliometers by comparison to a reference pyrheliometer', and 'Pyrheliometer Calibration', Chapter 7, Section 7.2.1.4 in the 'Guide to Instruments and Methods of Observation 2018' published by the World Meteorological Organisation (WMO).

### Calibration Procedure

- a) The PMO6 and MS-57 reference Pyrheliometer units are mounted on an STR-22G Sun Tracker and aligned normally with the sun.
- b) (PMO6) Measurement interval: 90 sec. (Dark phase: 45sec. including stabilizing time 5sec. / Open phase: 45sec. including stabilizing time 5sec.)
- c) (Reference Pyrheliometers) Sampling Interval: 1 sec. (averaging time: 5sec.)
- d) Number of data points: 40 stable measurement data points or more.
- e) The measurement values obtained by the PMO6 ( $W/m^2$ ) and the Reference pyrheliometers (output signal:  $\mu V$  / sensor temperature:  $^{\circ}C$ ) are logged with precise time stamps. The reference pyrheliometers' output signal is temperature corrected based on the measured sensor temperature and known temperature characteristics.
- f) The reference Pyrheliometer units' calibration figures are calculated from the measurement values of the PMO6 according to the measurement data criteria.
- g) Environmental conditions are applied as defined in appendix 4.





## Measurement Data Criteria

1. Standard deviation: Apply data less than 0.5% at  $2\sigma$  normal distribution
2. Measurement time: Apply data taken over 4 hours, centred at the culmination time
3. Solar irradiance conditions: Apply measurement data exceeding  $700\text{W}/\text{m}^2$

## Calculation of Sensitivity Value

- a) Acquire the signal output  $V$  ( $\mu\text{V}$ ) produced by the reference pyrhelimeter at the same time as the PMO6 irradiance, and apply temperature correction 'Tcoeff' (%), Relative ratio against the output at  $20^\circ\text{C}$ ) to each signal output.
- b) The sensitivity value of reference pyrhelimeter  $S$  ( $\mu\text{V}/\text{W}/\text{m}^2$ ) is obtained by dividing each signal output  $V$  ( $\mu\text{V}$ ) by the solar irradiance  $E_{\downarrow\text{Cavity}}$  ( $\text{W}/\text{m}^2$ ), measured simultaneously by the PMO6 and calculating the average value.

$$S = \frac{1}{n} \sum_{i=1}^n f \left( \frac{V T_{coeff}}{E_{\downarrow\text{cavity}}} \right)$$



## Calibrating a Field Pyrheliometer

Field pyrheliometers are calibrated by comparison against the reference pyrheliometer outdoors, according to the 'ISO 9059:1990 Solar energy -- Calibration of field pyrheliometers by comparison to a reference pyrheliometer' and 'Calibration of Pyrheliometers' as stated under Chapter 7, Section 7.2.1.4 on the 'Guide to Meteorological Instruments and Methods of Observation 2008' issued by the WMO.

### Calibration Procedure

- a) The reference pyrheliometer unit and the field pyrheliometer are mounted on the STR-22G Sun Tracker and aligned directly with the sun.
- b) Sampling Interval: 200msec. (averaging time 15sec.)
- c) Number of data points: 400 stable measurement data points or more.
- d) The field pyrheliometer and the reference pyrheliometer unit's measurement values are corrected based on the temperature characteristics of the detector.
- e) The voltage output ( $\mu\text{V}$ ) of the field pyrheliometer is averaged and matched with the data points (>400) of the reference Pyrheliometer ( $\mu\text{V}$ ,  $V_{\text{avg}}$ ). To obtain the sensitivity figure  $S$  ( $\mu\text{V}/\text{W}/\text{m}^2$ ) is multiplied by the temperature coefficient ( $T_{\text{coeff}}$ ) and divided by the direct solar irradiance ( $\text{W}/\text{m}^2$ ,  $E_{\downarrow \text{ref}}$ ) measured by the reference pyrheliometer.

$$S = \frac{1}{n} \sum_{i=1}^n f \left( \frac{V T_{\text{coeff}}}{E_{\downarrow \text{ref}}} \right)$$

## Reference Pyranometer Unit

EKO pyranometer models MS-80, MS-802, MS-60 and MS-40 are used as a reference for the indoor calibration of pyranometers per 'ISO 9846:1993 Solar Energy – Calibration of a pyranometer using a pyr heliometer' and the method defined in 'Alternate calibration using a pyr heliometer ("A new method for calibrating reference and field pyranometers" (Forgan, 1996)', Chapter 7, Section 3.1.3 of the 'Guide to Instruments and Methods of Observation 2018' published by the World Meteorological Organisation (WMO).

## Calibration Procedure

- a) The PMO6 and 2 reference pyranometer units of the same model are mounted on the STR-22G Sun Tracker and aligned to face direct sunlight.
- b) One reference pyranometer is shaded by a shading disk, or ball, attached to the Sun Tracker's arm to measure diffuse normal solar irradiance, while the other reference pyranometer remains unshaded to measure global normal solar irradiance. Each reference pyranometer alternates between being shaded and un-shaded near and at the solar noon.
- c) (PMO6) Measurement interval: every 90 sec. (Dark phase: 45sec including stabilizing time 5sec / Open phase: 45 sec. including stabilizing time 5sec.)
- d) (Reference Pyranometers) Sampling interval: 1 sec. (averaging time: 5sec.).
- e) Number of data: 40 stable measurement data points or more.
- f) The measurement values obtained by the PMO6 ( $W/m^2$ ) and reference Pyranometers (signal output:  $\mu V$  and sensor temperature:  $^{\circ}C$ ) are logged with precise time stamps. The pyranometer output signals will be temperature corrected based on the measured sensor temperature and known temperature characteristics.
- g) The sensitivity figures ( $\mu V / W/m^2$ ) of the two reference pyranometers can be calculated from the direct solar irradiance ( $W/m^2$ ) measured by the PMO6 and the measured voltage ( $\mu V$ ) of each reference pyranometer unit according to the measurement data criteria in the next paragraph.



## Measurement Data Criteria

1. Standard deviation: Apply data less than 0.5% at  $2\sigma$  normal distribution
2. Measurement time: Apply data taken over 4 hours, centred at the culmination time
3. Solar irradiance conditions: Apply measurement data exceeding  $700\text{W}/\text{m}^2$

## Calculation of Sensitivity Value

The sensitivity values of 2 reference pyranometers are calculated using an equation, and based on data gathered simultaneously under the same solar elevation angle, with reference irradiance measured by PMO6, diffuse solar irradiance, and global normal surface solar irradiance measured by 2 pyranometers.

The following equation (a) shows the signal output  $V_{dA}$  ( $\mu\text{V}$ ) of the diffuse solar irradiance measured by the reference pyranometer "A" and the signal output  $V_{gB}$  ( $\mu\text{V}$ ) of the inclined surface solar irradiance measured by the reference pyranometer "B" and direct solar irradiance  $E_{di}$  ( $\text{W}/\text{m}^2$ ) measured by PMO6 before the culmination time ( $t_0$ ).

Equation (b) shows the signal output  $V_{dB}$  ( $\mu\text{V}$ ) of the diffuse solar irradiance measured by the reference pyranometer "B" and the signal output  $V_{gA}$  ( $\mu\text{V}$ ) of the global normal surface solar irradiance measured by reference pyranometer "A" and direct solar irradiance  $E_{dir}$  measured by PMO6 after the culmination time ( $t_1$ ).

$R_A$  and  $R_B$  in equations (a) and (b) indicate the sensitivity value ( $\mu\text{V}/\text{W}/\text{m}^2$ ) of the 2 reference pyranometers.  $R_A$  and  $R_B$  are obtained using simultaneous formula (a) and (b) by assuming that the sensitivity constants are the same even when the diffuse solar irradiance and the global normal surface solar irradiance are measured.

- a) Extract the signal output  $V_A$ ,  $V_B$  ( $\mu\text{V}$ ) measured by the reference pyranometers at the same time as the PMO6, and apply temperature correction  $T_{coeff}$  (%), Relative ratio against the output at  $20^\circ\text{C}$ ) to each signal output. Pair the data measured under the same solar elevation angle.
- b)  $R_A$  and  $R_B$  can be obtained by calculating the sensitivity constant of the pyranometer standard from the data of each pair using the following equations (c) and (d) and then calculating the average value.

$$\frac{V_{gB}(t_0)}{R_B} = E_{dir}(t_0) + \frac{V_{dA}(t_0)}{R_A} \quad \dots(a)$$

$$\frac{V_{gA}(t_1)}{R_A} = E_{dir}(t_1) + \frac{V_{dB}(t_1)}{R_B} \quad \dots(b)$$

$$R_B = \frac{V_{dA}(t_0) V_{dB}(t_1) - V_{gB}(t_0) V_{gA}(t_1)}{V_{dA}(t_0) E_{dir}(t_1) - V_{gA}(t_1) E_{dir}(t_0)} \quad \dots(c)$$

$$R_A = \frac{V_{dA}(t_0) V_{dB}(t_1) - V_{gB}(t_0) V_{gA}(t_1)}{R_B E_{dir}(t_0) + V_{gB}(t_0)} \quad \dots(d)$$

## Calibrating a Field Pyranometer

EKO's field pyranometers are calibrated indoors according to the method defined in 'ISO 9847:1992 Solar energy -- Calibration of field pyranometers by comparison to a reference pyranometer (type II c)'.

Field pyranometers are calibrated against reference pyranometers of the same model using a AAA class Solar Simulator (1000W/m<sup>2</sup>, AM1.5G, +/-0.5% temporal stability) with an automated shutter.

The indoor calibration environment is controlled with an ambient temperature range of 25± 3°C.

### Calibration Procedure

- a) Before calibration can be performed, the solar simulator requires 30 minutes or longer to stabilize.
- b) The reference and field pyranometers are installed in a horizontal, automatic, 2-dimensional optical stage with high positioning precision, controlled by a PC with dedicated calibration software.
- c) A few minutes after the stabilization time, the offset output  $V_{dref}$  (μV) and  $V_{dtest}$  (μV) are measured in the dark.
- d) The automatic stage moves the reference pyranometer to a specific position under the solar simulator. The sensor entrance optics remains aligned with the radiation source; the shutter opens automatically, and the irradiation starts.
- e) After irradiating for the time specified for each pyranometer model, record 10 points of the signal output of the reference pyranometers  $V_{iref}$  (μV).

- f) The automatic stage then moves the field pyranometer to alternate it with the reference pyranometer and places the field pyranometer under the same irradiation conditions.
- g) After irradiating for the same amount of time as the reference pyranometer, record 10 points of the signal output of the field pyranometers  $V_{itest}$  ( $\mu V$ ).
- h) The automatic stage moves the reference pyranometer back to the same position, alternates it with the field pyranometer, and exposes it again to the radiation source.
- i) After irradiating for the same specified time, record 10 points of the signal output of the reference pyranometers  $V_{iref}$  ( $\mu V$ ).
- j) After the measurement process is completed, the software automatically determines whether the standard deviation of the signal output of the reference pyranometer  $V_{iref}$ , the 20 points recorded before and after the measurement of the product, meets the acceptance criteria.
- k) Calculate the average solar irradiance  $E_{\downarrow ref}$  by subtracting the offset output  $V_{dref}$  in the dark state from the measurement data of the reference pyranometer  $V_{iref}$  and dividing by the sensitivity constant  $S_{ref}$ .
- l) The sensitivity value of the field pyranometer  $S$  is calculated by subtracting the output  $V_{dtest}$  in the dark state from the output at the time of irradiation  $V_{itest}$  and dividing it by the solar irradiance  $E_{\downarrow ref}$  obtained by the reference pyranometer.

$$E_{\downarrow ref} = \frac{1}{20} \sum_{i=1}^{20} f \left( \frac{V_{iref} - V_{dref}}{S_{ref}} \right)$$

$$S = \frac{1}{10} \sum_{i=1}^{10} f \left( \frac{V_{itest} - V_{dtest}}{E_{\downarrow ref}} \right)$$

## APPENDIX

### 1. Scope of ISO/IEC 17025:2017 Calibrations

The scope of accreditation (certificate ref: L19-266 / www.pjllabs.com) specifies the Calibration and Measurement Capability (CMC) for the applicable sensor categories, calibration methods and measurement uncertainty with a confidence level of 95% ( $k=1.96$ ).

The CMC states that the accredited processes represent the smallest measurement uncertainties attainable by the laboratory when performing a nearly ideal device's routine calibration under nearly ideal conditions. The actual measurement uncertainty associated with a specific calibration performed will typically be larger than the CMC specified for the same calibration since the device's capability and performance, and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.

#### Sample

EKO Instruments' Certificate of Accreditation' ISO17025 (published in 2019): Calibration scope and Uncertainty of Calibration in EKO Instrument calibration laboratory

|   <b>Certificate of Accreditation: Supplement</b><br><b>EKO Instruments Co., Ltd.</b><br>1-21-8 Hatagaya Shibuya-ku, Tokyo 151-0072<br>Contact Name: Shuji Yoshida Phone: 03-3469-6711 |  |  |  |
|--|--|--|--|
| Accreditation is granted to the facility to perform the following calibrations:  |  |  |  |
| Optical  |  |  |  |
| MEASURED INSTRUMENT, QUANTITY OR GAUGE   | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE    | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (%) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED   |
| Pyrheliometer (outdoor calibration) ISO9060 1st Class <sup>g</sup>   | 700 W/m <sup>2</sup> to 1 200 W/m <sup>2</sup> | 0.59 % of reading  | Pyrheliometer and Pyranometer Calibration Operating Instructions (Clause 2) (LM-10)<br>On basis of:<br>WMO-No.8:2008 and ISO9059:1990<br><br>Standard Pyrheliometer (MS-57)<br>Data logger (Cadac 21)                |
| Pyrheliometer (indoor calibration) ISO9060 1st Class <sup>f</sup>  | 700 W/m <sup>2</sup> to 1 200 W/m <sup>2</sup> | 0.58 % of reading  | Pyrheliometer Indoor Calibration Operating Instructions (Clause 2) (LM-10A)<br>On basis of:<br>WMO-No.8:2008 and ISO9847:1992<br><br>Standard Pyrheliometer (MS-57)<br>Digital multi-meter (34401A)                  |
| Pyranometer ISO9060 Secondary Standard <sup>g</sup>  | 700 W/m <sup>2</sup> to 1 400 W/m <sup>2</sup> | 0.59 % of reading  | Pyrheliometer and Pyranometer Calibration Operating Instructions (Clause 3) (LM-10)<br>On basis of:<br>WMO-No.8:2008 and ISO9847:1992<br><br>Standard pyranometer (MS-802 and MS-80)<br>Digital multi-meter (34401A) |
| Pyranometer ISO9060 1st Class <sup>g</sup>   |  | 0.64 % of reading  |  |
| Pyranometer ISO9060 2nd Class <sup>g</sup>   |  | 0.64 % of reading  |  |
| Silicon-pyranometer (ML-01), (ML-02) <sup>g</sup>  | 700 W/m <sup>2</sup> to 1 400 W/m <sup>2</sup> | 1.8 % of reading   | Pyrheliometer and Pyranometer Calibration Operating Instructions (Clause 3) (LM-10)<br>On basis of:<br>WMO-No.8:2008 and ISO9847:1992<br><br>Standard silicon-pyranometer (ML-01)<br>Digital multi-meter (34401A)    |

Issue: 06/2019

This supplement is in conjunction with certificate #L19-266

Page 3 of 4

## 2. Solar Irradiance Definitions

Solar Irradiance  $E_{\downarrow}$  is the rate at which radiant energy from the sun is incident on a surface per unit area of surface, commonly expressed in  $[W/m^2]$ .

### a) Global Solar Irradiance ( $E_{g\downarrow}$ )

The power density incident on a surface per unit area from the hemispheric sky (hemisphere:  $2\pi$  steradian). The rate at which solar irradiance is received directly from the sun and from the atmosphere on the earth's surface. It is equal to the sum of the direct solar irradiance and the diffuse solar irradiance.


### b) Direct Solar Irradiance ( $E$ )

The power density incident on a surface per unit area from a small solid angle of solar radiation directly from the sun disk.

Reference Solar Spectrum: For clear sky condition a standard irradiance spectrum is defined by ISO 9845 (Air mass 1.5 solar spectrum / Direct normal incidence up to  $5.8^\circ$  FOV / hemispherical radiation on an equator-facing,  $37^\circ$  tilted plane for an albedo of 0.2 / spectral range 285 – 4000 nm).

### c) Diffuse Solar Irradiance ( $E_{d\downarrow}$ )

The rate at which solar radiation reaches the earth's surface per unit area after it is scattered by molecules, aerosols, clouds and other particles in the air.



### 3. Measurement Devices

The specifications and classifications of pyrhemometers and pyranometers are prescribed in 'ISO 9060:2018, Specification and classification of instruments for measuring hemispherical solar and direct solar radiation'.

#### a) Pyrhemometer

Pyrhemometer designed for measuring direct normal irradiance with a limited Field of View (FOV) of  $\angle 5^\circ$  (opening half-angle of  $\angle 2.5^\circ$ ). The captured solar energy is absorbed in the wavelength range 0.3~3.0 $\mu\text{m}$ .

#### b) Pyranometer

Pyranometer designed for measuring global irradiance on a plane receiver surface in the wavelength range of 0.3~3.0 $\mu\text{m}$ .

#### c) Sun Tracker

A Sun Tracker is an automated device to track the sun with high angular precision (slope angle allowance less than  $0.25^\circ$ ). It is based on a dual-axis motion control system, which can rotate in zenith and azimuth directions. EKO uses the STR-22G Sun Tracker, which has a higher tracking accuracy of less than  $0.01^\circ$ .

#### d) Data Acquisition System

A digital multi-meter or data-logger, capable of recording and collecting data from pyrhemometers and pyranometers. The resolution of the device used for calibration is less than 0.05%.

#### 4. Criteria and Environmental Conditions

The following environmental conditions and criteria apply to the outdoor calibration of the reference pyranometer and reference pyrhelimeter.

##### a) Direct Solar Irradiance ( $E_g \downarrow$ )

Direct solar irradiance (measured by the Absolute Cavity Radiometer) exceeding  $700\text{W/m}^2$ . Ratio (direct solar irradiance intensity / global solar irradiance intensity) is greater than 0.85.

##### b) Environmental Parameter 1 (Wind Speed)


Low wind speed conditions are preferable. Measurements in the case of wind speeds exceeding  $> 5.0\text{m/s}$  must be avoided, especially when the wind direction is within  $\pm 30^\circ$  of the sun's direction. The wind could potentially blow inside the Absolute Cavity Radiometer and may cause larger uncertainties.

##### c) Environmental Parameter 2 (Ambient Temperature)

The ambient temperature should be within the specified operating temperature of the Absolute Cavity Radiometer, reference pyrhelimeter and reference pyranometer.

##### d) Environmental Parameter 3 (Atmospheric Conditions)

Only days with few clouds and stable atmospheric conditions allow for outdoor calibration (clouds amount must be less than 12.5% in order to take stable measurements). No clouds are accepted within the sensor field of view of  $15^\circ$  during the period of measurements. In cases where clouds cross this range during the measurement series, record its time period and remove the measurement data.



## 5. Calibration Certificates

A calibration certificate is produced for each pyranometer or pyrhemliometer and contains the following information.

- a) Identification of the calibration
- b) Calibration conditions
- c) Calibration results
- d) Reference instruments
- e) Calibration procedure
- f) Calibration uncertainties
- g) Traceability
- h) Authorisation of the calibration and report

### Sample

#### Calibration Certificate for Pyranometer

|   |   |   |                 |
|---|---|---|-----------------|
|   |   | <small>EKO INSTRUMENTS CO., LTD.<br/>1-21-8 Hatagaya, Shibuya-ku,<br/>Tokyo 151-8572, Japan<br/>P: +81 3 3469 6713<br/>F: +81 3 3469 6719<br/>www.eko.co.jp</small> |                 |
| <h3>Calibration Certificate</h3>  |   |   |                 |
| Requester :   | Customer info (name)<br>Customer info (address) | <b>ISO/IEC 17025:2017</b><br>   |                 |
| Manufacturer :  | EKO Instruments Co., Ltd.<br>Pyranometer        | Certificate Number :  | Sxxxxxxx-xx-xxx |
| Description :   | MS-xx   | Issue Date :  | MM DD, YYYY     |
| ISO Classification :  | Class x   | Calibration Date :  | MM DD, YYYY     |
| Serial Number :   | Sxxxxxxx  | Calibration Procedure :   | LM-10           |
| Accessories :   | Cable   |   |                 |
| <b>Calibration Conditions</b><br>Temperature xx±x [°C]<br>Irradiance xxx±xxx [W/m <sup>2</sup> ]<br>Solar Simulator Instability x.x [%]   |   |   |                 |
| <b>Calibration Results</b><br>Sensitivity xx.xx [μV/W·m <sup>2</sup> ]<br>Uncertainty x.xx [%] (Coverage factor k = 1.96)   |   |   |                 |
| The above product is calibrated and traceable to the reference pyranometer in compliance with ISO9847 Direct beam calibration (type IIC). Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM). |   |   |                 |
| Reference Instruments   | Model   | S/N   | Calibration due |
| Pyranometer   | MS-xx   | xxxxxxxx  | YYYY/MM/DD      |
| Digital Multimeter  | xxxxx   | xxxxxxxx  | YYYY/MM/DD      |
| EKO Instruments Co., Ltd.<br>1-21-8, Hatagaya, Shibuya-ku, Tokyo, 151-0072, Japan<br>M.Kita / Calibration responsible<br><br>Certificate Number : Sxxxxxxx-xx-xxx   |   |   |                 |
|   |   | Page 1/2  |                 |

| Calibration Results |                                     |                         |                                    |    |                                     |                         |                                    |
|---------------------|-------------------------------------|-------------------------|------------------------------------|----|-------------------------------------|-------------------------|------------------------------------|
| n                   | Ref. Irradiance [W/m <sup>2</sup> ] | Pyranometer Output [mV] | Sensitivity [μV/W·m <sup>2</sup> ] | n  | Ref. Irradiance [W/m <sup>2</sup> ] | Pyranometer Output [mV] | Sensitivity [μV/W·m <sup>2</sup> ] |
| 1                   | xxxx.x                              | x.xxx                   | x.xxx                              | 6  | xxxx.x                              | x.xxx                   | x.xxx                              |
| 2                   | xxxx.x                              | x.xxx                   | x.xxx                              | 7  | xxxx.x                              | x.xxx                   | x.xxx                              |
| 3                   | xxxx.x                              | x.xxx                   | x.xxx                              | 8  | xxxx.x                              | x.xxx                   | x.xxx                              |
| 4                   | xxxx.x                              | x.xxx                   | x.xxx                              | 9  | xxxx.x                              | x.xxx                   | x.xxx                              |
| 5                   | xxxx.x                              | x.xxx                   | x.xxx                              | 10 | xxxx.x                              | x.xxx                   | x.xxx                              |
| Σ Mean (n= 10)      |                                     |                         | x.xxx                              |    |                                     |                         | x.xxx                              |
| Std. deviation      |                                     |                         | x.xxx                              |    |                                     |                         | x.xxx                              |

| Calibration Procedure   |  |
|---|--|
| The pyranometer was calibrated against a calibrated reference pyranometer using a 1000 W/m <sup>2</sup> (AM1.5 class AAA) sun simulator as source. The pyranometer is situated on a horizontal table and aligned to the optical axis of a normal incidence light source. By alternating the position of the calibrated reference pyranometer with the test pyranometer the output signal of both pyranometers are recorded and used to solve the equation of the unknown sensitivity variable. The operating conditions are maintained constant (e.g. ambient temperature and normal incidence irradiance), hence the pyranometer uncertainty figure from 10 readings (n) is determined by taking into account the sensitivity uncertainty (Us), uncertainty of temperature (Ut), normal distribution (d), and repeatability (Ur) of the reference pyranometer and the max. deviation of the incident irradiance between the measurement intervals. |  |

| Uncertainty   |                                  |
|---|----------------------------------|
| The expanded calibration uncertainty associated with the pyranometer sensitivity figure is calculated as the square root of the sum of the squares of the reported uncertainties: |                                  |
| 1) Uncertainty of reference (Us)  | x.xx%                            |
| 2) Uncertainty of temperature (Ut)  | x.xx%                            |
| 3) Repeatability (Ur)   | x.xx%                            |
| 4) Distribution (d)   | x.xx%                            |
| Total uncertainty (Uc)  | x.xx%                            |
| Expanded uncertainty (U)  | x.xx% (Coverage factor k = 1.96) |

Expanded Uncertainty = 1.96 x √((Us<sup>2</sup> + Ut<sup>2</sup> + Ur<sup>2</sup> + d<sup>2</sup>))

| Traceability   |  |
|--|--|
| Every 1 year, the reference pyranometer MS-80 is calibrated against the primary standard PMO-6 according to the sun-and-shade method under ISO9846. The primary standard is directly traceable to the WRR (World Radiometric Reference) and maintained in the group of standard radiometers calibrated every 5 years during the IPC. The data logger is traceable to JEMIC (Japan Electric Meters Inspection Corporation). |  |

Certificate Number : Sxxxxxxx-xx-xxx

Page 2/2



## 6. Measurement Uncertainty

The uncertainty of measurements is defined for each type of measurement and sensor model. According to the 'Guide of the Expression of Uncertainty in Measurement' (GUM), all parameters that contribute to measurement uncertainty are identified and taken into account using the standard analysis methods.

### Sample

Budget Sheet: Calculated Uncertainty value of Class A pyranometer product (Rev. 2019)

| Type | Symbol | Uncertainty factor            | Division              | Value (±)     | Unit                  | Distribution            | Divisor    | Standard Uncertainty | Squared value  | Remarks  |
|------|--------|-------------------------------|-----------------------|---------------|-----------------------|-------------------------|------------|----------------------|----------------|--|
| B    | STD    | Reference unit                | Reference Pyranometer | <b>0.567</b>  | %                     | Std. deviation          | 1.96       | <b>0.28</b>          | <b>0.0802</b>  | Reference unit (Including measurement variation)   |
| A    | GSI    | Variation of measurement data | Pyranometer           | <b>0.008</b>  | %                     | ---                     | 1          | <b>0.008</b>         | <b>0.00006</b> | Uncertainty of the mean value of the sensitivity constant  |
| A    | RP     | Repeatability                 | Pyranometer           | <b>0.02</b>   | %                     | ---                     | 1          | <b>0.02</b>          | <b>0.00028</b> | Measurement repeatability (Including light source stability)   |
| B    | Td     | Temperature response          | Pyranometer           | <b>0.12</b>   | %                     | Rectangular             | $\sqrt{3}$ | <b>0.069</b>         | <b>0.0048</b>  | TC response error due to the change of the controlled room temperature at $25 \pm 3^\circ\text{C}$ ( $6^\circ\text{C}$ ) |
| B    | DM     | Measurement uncertainty       | Digital multimeter    | <b>0.0075</b> | %                     | Rectangular             | $\sqrt{3}$ | ---                  | ---            | Specification value (Ignored: Too small value and has no effect)   |
| B    | NL     | Non-linearity                 | Pyranometer           | <b>0.2</b>    | %                     | Rectangular             | $\sqrt{3}$ | ---                  | ---            | Specification value (Excluded: Due to irradiate with a constant value around $1000\text{W}/\text{m}^2$ )                 |
| B    | Z01    | Zero offset A                 | Pyranometer           | <b>7</b>      | $\text{W}/\text{m}^2$ | Rectangular             | $\sqrt{3}$ | ---                  | ---            | Actual value (Omitted the offset when measured in the dark)  |
| B    | Z02    | Zero offset B                 | Pyranometer           | <b>2</b>      | $\text{W}/\text{m}^2$ | Rectangular             | $\sqrt{3}$ | ---                  | ---            | Actual value (Excluded: Due to calibrate in a constant temperature environment)  |
| B    | CR     | Cosine response               | Pyranometer           | <b>10</b>     | $\text{W}/\text{m}^2$ | Rectangular             | $\sqrt{3}$ | ---                  | ---            | Specification value (Excluded: Due to irradiate from zenith angle $0^\circ$ only)  |
| B    | TI     | Tilt response                 | Pyranometer           | <b>0.17</b>   | %                     | Rectangular             | $\sqrt{3}$ | ---                  | ---            | Specification value (Excluded: Due to place and irradiate on a horizontal plane)   |
|      | uc( )  | Combined Uncertainty          |                       |               | %                     | Std. deviation          |            | <b>0.292</b>         | <b>0.085</b>   |  |
|      | U      | Expanded Uncertainty          |                       |               | %                     | Std. deviation (k=1.96) |            | <b>0.584</b>         |                |  |



## 7. Sensor Characteristics

Measurement results are provided for each class A pyranometer and class A pyrhemliometer.

### Sample

Fig1. (Left) Temperature Response Measurement Test Report for Pyranometer

Fig2. (Right) Directional Response Measurement Report for Pyranometer

